

**DENTAL X-RAY**

# **BELRAY II 097**


## **OPERATOR'S INSTRUCTIONS**

- *Wall Mount Type* ..... WK
- *Floor Mount Type* ..... FK1/FK2
- *Mobile Type* ..... FM
- *Room Mount Type* ..... RK
- *Ceiling Mount Type* ..... CK



**⚠ WARNING**

This X-ray unit may be dangerous to patient and operator unless safe exposure factors, operating instructions and maintenance schedules are observed.

 **Belmont<sup>®</sup>**



## INDEX

[1] INTRODUCTION.....	1
[2] MAJOR COMPONENTS.....	3
[3] FUNCTION OF CONTROLS .....	6
[4] OPERATING PROCEDURES .....	9
[5] ESTIMATED AIR KERMA.....	9
[6] HAND EXPOSURE SWITCH.....	9
[7] DIGITAL IMAGING SYSTEM .....	10
[8] INFECTION CONTROL AND CLEANING .....	10
[9] ERROR CODES .....	11
[10] MAINTENANCE.....	12
[11] DISPOSAL .....	12
[12] TECHNICAL DATA.....	14
[13] ELECTROMAGNETIC COMPATIBILITY (EMC) .....	16
[14] LABEL LOCATION.....	18
[15] OTHER INFORMATION.....	20

# [1] INTRODUCTION

## 1. GENERAL

This manual provides information for the operation and maintenance procedures and technical specifications for BELRAY II 097 dental x-ray. The instructions contained in this book should be thoroughly read and understood before operation. BELRAY II 097 has no user serviceable items. Repair should be performed by qualified dealer service personnel. Any part of this x-ray unit shall not be maintained or serviced while in use with a patient.

## 2. INTENDED PURPOSE

- a. BELRAY II 097 is an extra-oral source dental radiographic x-ray unit. This unit is an active device intended to generate and control diagnostic purpose ionizing radiation. The absorption pattern of x-ray beam recorded on intraoral image receptor is used for general-purpose, routine, dental radiography examinations of diseases of the teeth, jaw and oral cavity structures.
- b. Patient Population: All patients with oral diseases are eligible, without distinction as to age, gender, etc. However, infants/children who are unable to remain still during imaging or who are unable to hold the film, sensor, or other imaging device in their mouths are excluded from the scope.
- c. Intended Users: Qualified Healthcare Professionals trained in the use of the device, familiar with the use and application of x-ray imaging systems and the local x-ray protection requirements.

## 3. COMPONENTS IDENTIFICATION OF X-RAY SYSTEM "BELRAY II 097"

- a. Tube housing assembly: 097-H
- b. X-ray controls: 097-CM (main controller), 097-CS (sub controller)
- c. Cones: 097-R (regular), 097-L (long)
- d. Collimator: 097-REC (rectangular)
- e. Balance arm: 097-A
- f. RK stand: 097-RK

## 4. DECLARATION OF CONFORMITY

We declare BELRAY II 097 x-ray unit complies with following regulation and directive.

MDR (Medical Device Regulation): Regulation (EU) 2017/745 Annex II and III

RoHS Directive: 2011/65/EU category 8 of Annex I

## 5. CLASSIFICATION

5-1. According to Medical Device Regulation, BELRAY II 097 is classified as CLASS IIb medical device by the rule 10 of MDR ANNEX VIII.

5-2. According to IEC60601-1, BELRAY II 097 is classified as follows.

- a. Protection against electric shock: Class I Equipment
- b. Type of applied parts: Type B (RK type only)
- c. Protection against ingress of water: Ordinary
- d. Mode of operation: Non continuous (Duty Cycle = 1:50,  
Max. ON time: 3.2 sec, Min. OFF time: 12 sec.)
- e. Equipment not suitable for use in the presence of a flammable anesthetic mixture with air or with oxygen or nitrous oxide.
















































## 6. NOTICE TO THE USER

- a. This x-ray unit may be dangerous to patient and operator unless safe exposure factors, operating instructions and maintenance schedules are observed.
- b. Only qualified and authorized personnel may operate this equipment observing all laws and regulations concerning protection. The operator must:
  - have means for audio and visual communication with the patient.
  - have full view of kV, mA, timer selections and exposure warning light.
  - be at least 2 m away from the x-ray head and patient and out of the path of the x-ray beam or be positioned behind a protective device.
  - fully use all radiation protection devices, accessories and procedures available to protect the patient and operator from x-ray radiation.
- c. Any serious incident occurred in relation to the device should be reported to the manufacturer and the competent authority of the Member State in which the user is established.

- d. This x-ray unit must be properly managed and used to meet the requirements of EU Directive 2013/59/EURATOM. For example, the annual radiation dose should be kept below the limit specified by the Directive. If the local laws or regulations are stricter than the Directive, the user shall comply with those laws or regulations.


## 7. SYMBOL

In this book, on the labels or on the control panel of BELRAY II 097, following symbols are used. Confirm the meanings of each symbol by the table below.

	Follow Instructions for use		Protection against electric shock: Type B		ON (POWER)		OFF (POWER)
	Protection Grounding		Exposure Switch		X-ray Emission		Ready
	Upper Incisor		Upper Cuspid & Pre Molar		Upper Molar		Occlusal
	Lower Incisor		Lower Cuspid & Pre Molar		Lower Molar & Bite Wing		Bite Wing
	Digital Imaging		Patient Child		Patient Adult		Patient Large Adult
	Regular Cone		Long Cone		Authorized Representative in The European Community		Weight of whole unit
	Manufacturer		Date of Manufacture		Serial Number		Catalogue Number
	Separate Collection for Electrical and Electronic Equipment		Compliance with European directive required		Medical device		Type of the device
	Identification for each component		Rated input to the device		Rated output from the device		Max. ON time: 2 second. Min. OFF time: 12 second
	Manufactured in Japan (used with date symbol)		Brand symbol of Takara Belmont group		Unique Device Identifier		Electronic instructions for use
	X-ray tube		Source to skin distance			FOCAL SPOT VALUE INHERENT FILTRATION ADDED FILTRATION TOTAL FILTRATION RADIATION LEAKAGE RATE	
	X-ray field size at cone end						
		Power supply requirements Rated Voltage [Vac] Max Apparent Resistance [Ω] Over Current Release [A]				CAUTION DO NOT MOVE ENTIRE X-RAY UNIT WITH ARM EXTENDED.	
		CAUTION ! DO NOT RELEASE THIS BAND UNTIL X-RAY HEAD IS INSTALLED.				WARNING Keep casters in the lock position, unless moving the equipment. To avoid injury, do not push or lean on the equipment.	

## [2] MAJOR COMPONENTS


### 1. WALL MOUNT TYPE (WK)



- ① Main Power Switch
- ② X-Ray Head
- ③ Cone
- ④ Yoke
- ⑤ Arm Collar
- ⑥ Balance Arm
- ⑦ Horizontal Arm
- ⑧ Main Controller
- ⑨ Sub Controller
- ⑩ Hand Exposure Switch (Option)

Fig.2-4 Major Components for WK

### 2. FLOOR MOUNT TYPE (FK1/FK2)



- ① Main Power Switch
- ② X-Ray Head
- ③ Cone
- ④ Yoke
- ⑤ Arm Collar
- ⑥ Balance Arm
- ⑦ Horizontal Arm (300 mm)
- ⑧ Pole Bush
- ⑨ Back Supporter
- ⑩ Pole
- ⑪ Base Cover
- ⑫ Mounting Plate
- ⑬ Main Controller
- ⑭ Sub Controller
- ⑮ Hand Exposure Switch

Type FK1

Type FK2

Fig.2-1 Major Components for FK1/FK2

### 3. MOBILE TYPE (FM)

- ① Main Power Switch
- ② X-Ray Head
- ③ Cone
- ④ Yoke
- ⑤ Arm Collar
- ⑥ Balance Arm
- ⑦ Pole Bush
- ⑧ Pole
- ⑨ Pole Base
- ⑩ Leg Bar (long)
- ⑪ Leg Bar (Short)
- ⑫ Lock Caster
- ⑬ Standard Caster
- ⑭ Main Controller
- ⑮ Sub Controller
- ⑯ Hand Exposure Switch

#### WARNING

Keep casters in the lock position, unless moving the equipment. To avoid injury, do not push or lean on the equipment.

#### CAUTION

Do not move entire x-ray unit with arm extended.




Fig.2-2 Major Components for FM

### 4. ROOM MOUNT TYPE (RK)




Fig.2-3 Major Components for RK

- ① Main Power Switch
- ② X-Ray Head
- ③ Cone
- ④ Yoke
- ⑤ Arm Collar
- ⑥ Swing Arm 1
- ⑦ Swing Arm 2
- ⑧ Sliding Post
- ⑨ Column Cover
- ⑩ Column
- ⑪ Backrest Cushion (applied part)
- ⑫ Seat (applied part)
- ⑬ Gas Cylinder
- ⑭ Base Plate
- ⑮ Main Controller
- ⑯ Sub Controller
- ⑰ Hand Exposure Switch (option)

## 5. CEILING MOUNT TYPE (CK)




Fig.2-5 Major Components for CK

## 6. SUB CONTROLLER




Fig.2-6 Sub Controller Switches

- |   |                                 |
|---|---------------------------------|
| ① Sub Controller Front Panel            | ⑩ Cone Type Selection Switch    |
| ② Ready Light                           | ⑪ Film Speed Selection Switch   |
| ③ Exposure Time Adjusting Switch (Down) | ⑫ Digital Imaging Switch        |
| ④ Exposure Time Adjusting Switch (Up)   | ⑬ 4 mA Selection Switch         |
| ⑤ Tooth Selection Switch (T1)           | ⑭ 7 mA Selection Switch         |
| ⑥ Tooth Selection Switch (T2)           | ⑮ Patient Size Selection Switch |
| ⑦ Tooth Selection Switch (T3)           | ⑯ Exposure Time Display Window  |
| ⑧ Tooth Selection Switch (T4)           | ⑰ Exposure Warning Light        |
| ⑨ Tooth Selection Switch (T5)           | ⑱ Exposure Switch               |



### [3] FUNCTION OF CONTROLS

#### ① Main Power Switch



Pushing the upper side of this switch to the ON position energizes the x-ray unit. (Ready light and pre-select lights for cone type, film or digital, 4 mA, 7 mA, and patient size illuminate.) It is recommended to keep this switch OFF when the unit is not in use, in order to prevent an accidental exposure.

**IMPORTANT: To prevent the risk of an accidental exposure, push the lower side of this switch to the OFF position, when the unit is not in use.**

#### ② Ready Light

This light illuminates when the line voltage is within operable range (207 - 253Vac). When this light is not on, exposure can not be made.

#### ③ ④ Exposure Time Adjusting Switches

By momentarily pushing the  (or ) switch, the exposure time displayed increases (or decreases) by one increment. By keeping the switch depressed more 2 sec., the exposure time displayed increases (or decreases) continuously until the switch is released.

Model 097 has the following 23 exposure time settings:

0.00, 0.02, 0.03, 0.04, 0.05, 0.06, 0.08, 0.10, 0.13, 0.16, 0.20, 0.25, 0.32, 0.40, 0.50, 0.63, 0.80, 1.00, 1.25, 1.60, 2.00, 2.50, 3.20 (sec)

#### ⑤ - ⑨ Tooth Selection Switches (T1-T5)

Pushing one of these switches sets the exposure time automatically for the following ⑩ - ⑮.

- ⑤ T1: Incisor of Mandible
- ⑥ T2: Incisor of Maxilla, Cuspid & Premolar of Mandible
- ⑦ T3: Cuspid & Premolar of Maxilla, Molars of Mandible, Bitewing
- ⑧ T4: Molar of Maxilla, Bitewing Molars
- ⑨ T5: Occlusal

If the T1 switch ⑤ is depressed more than 3 sec. unit goes into " Lock Mode". In lock mode, the only functional switch is the power switch. To exit from the lock mode, depress the T1 switch more than 3 sec. again.

#### ⑩ Cone Type Selection Switch

By depressing this switch for less than 2 sec., selected cone type can be confirmed. If round cone is selected, "rnd" is displayed. If the cone with rectangular collimator is selected, "rEC" is displayed. Depressing this switch more than 2 sec., changes the cone type by the following order.

Short cone (Round) → Short cone with rectangular collimator → Long cone (Round)  
Long cone with rectangular collimator → Short cone (Round) → continued.

#### ⑪ Film Speed Selection Switch



a. BELRAY II has 16 film speed settings. (F.00 - F.15)



Two speed settings are pre-set at the factory (a & b) and can be selected with switch ⑪.

a = Film speed No. F.09 (equivalent to ISO speed group "D", or Kodak Ultra-Speed film)

b = Film speed No. F.04 (equivalent to ISO speed group "F/E", or Kodak InSight film)

Including these two speeds, BELRAY II 097 x-ray can provide 16 different film speeds (F.00 - F.15) and any two of them can be programmed for easy selection. If doctor uses a different film speed, or prefers darker (or lighter) radiographs, the new speed can be programmed as follows. Higher speed settings make films darker. If film speed is increased by 1, exposure time becomes 25 % longer.



1. Keep the 4 mA selection switch and 7 mA selection switch depressed simultaneously for more than 3 seconds.  
Release the switches if the ready light starts to flash.
2. Push F switch momentarily until the "a" light above the F switch illuminates. The exposure time display window shows the present film speed for "a" setting. (The factory default setting, F.09 should be displayed.) By depressing  or  switch, increase or decrease film speed number until desired number for "a" setting is displayed.

3. To change the “b” setting from the factory default, F.04, push F switch momentarily until the “b” light illuminates. By depressing  or  switch, increase or decrease film speed until the desired number for “b” setting is displayed.
  4. Press T1 switch to store these settings, then turn the main power switch off.
- b. Pushing **Film Speed Selection Switch** ⑪ momentarily displays the selected film speed setting in the **Exposure Time Display Window** ⑩
- Depressing this switch for more than 2 sec. changes the film type being selected.
- c. If the **Digital Imaging Switch** ⑫ is depressed, both of the film speed indicating lights (a & b) are turned off.

### ⑫ Digital Imaging Switch

If a digital imaging system is used, shorter exposure time is often required. BELRAY II has 16 speeds for digital imaging (d.00 - d.15). Pushing this switch momentarily displays the speed being selected in the **Exposure Time Display Window** ⑩. With the factory speed setting d.08, the exposure time becomes half of F.08 setting.

As the sensitivity is different according to each manufacturer of digital imaging sensors, this setting should be adjusted. To get a darker image, increase the speed setting and to get a lighter image, decrease the speed setting. If the speed setting is increased by 1, exposure time becomes 12 % longer.

1. Keep 4 mA selection switch and 7 mA selection switch depressed simultaneously for more than 3 seconds.
2. Push D switch momentarily until the light above the D switch illuminates and the exposure time display window shows the present speed setting. (The factory default setting d.08 should be displayed.)
3. By depressing  or  switch, increase or decrease speed until the desired number is displayed.
4. Press **T1 switch** to store these settings, then turn the main power switch off.

### ⑬ 4 mA Selection Switch

By momentarily depressing this switch, the tube current is set at 4 mA.

When Film switch is depressed, the tube current setting will be automatically changed to 7 mA.

### ⑭ 7 mA Selection Switch

By momentarily depressing this switch, the tube current is set at 7 mA.

When digital switch is depressed, the tube current setting will be automatically changed to 4 mA.

**TABLE 1. Speed Setting and Exposure Time (Reguler Cone)**

[ unit : sec.]




Speed Setting	mA	Child					Adult					Large Adult				
		T1	T2	T3	T4	T5	T1	T2	T3	T4	T5	T1	T2	T3	T4	T5
F.09	4	0.20	0.32	0.40	0.50	0.80	0.32	0.50	0.63	0.80	1.25	0.40	0.63	0.80	1.00	1.60
	7	0.10	0.20	0.20	0.32	0.40	0.20	0.32	0.40	0.50	0.63	0.20	0.40	0.40	0.63	0.80
F.04	4	0.08	0.13	0.16	0.20	0.32	0.13	0.20	0.25	0.32	0.50	0.16	0.25	0.32	0.40	0.63
	7	0.05	0.08	0.10	0.13	0.16	0.08	0.13	0.16	0.20	0.32	0.10	0.16	0.20	0.25	0.32
d.08	4	0.08	0.13	0.16	0.20	0.32	0.13	0.20	0.25	0.32	0.50	0.16	0.25	0.32	0.40	0.63
	7	0.05	0.08	0.10	0.13	0.20	0.08	0.13	0.16	0.20	0.32	0.10	0.16	0.20	0.25	0.40

**TABLE 2. Speed Setting and Exposure Time (Long Cone)**

[ unit : sec.]

Speed Setting	mA	Child					Adult					Large Adult				
		T1	T2	T3	T4	T5	T1	T2	T3	T4	T5	T1	T2	T3	T4	T5
F.09	4	0.40	0.63	0.80	1.00	1.60	0.63	1.25	1.25	2.00	2.50	0.80	1.25	1.60	2.00	3.20
	7	0.25	0.40	0.50	0.63	1.00	0.40	0.63	0.80	1.00	1.60	0.50	0.80	1.00	1.25	2.00
F.04	4	0.16	0.25	0.32	0.50	0.63	0.25	0.50	0.50	0.80	1.00	0.32	0.50	0.63	1.00	1.25
	7	0.10	0.16	0.20	0.25	0.40	0.16	0.25	0.32	0.40	0.63	0.20	0.32	0.40	0.50	0.80
d.08	4	0.16	0.32	0.32	0.50	0.63	0.32	0.50	0.63	0.80	1.00	0.32	0.63	0.63	1.00	1.25
	7	0.10	0.16	0.20	0.25	0.40	0.16	0.25	0.32	0.40	0.63	0.20	0.32	0.40	0.50	0.80

### ⑮ Patient Size Selection Switch

This switch alters the selection of patient type/size to be radiographed (child → adult → large adult → child) and sets the exposure time automatically. If the weight of child is less than 20kg, press  switch once after setting to child. If the weight of child is over 30kg and less than 50kg, press  switch once after setting to child. If the weight of child is over 50kg and less than 70kg, press  switch twice after setting to child. If the weight of child is over 70kg, set to adult.

**NOTE: Setting or adjusting the exposure time manually (with  or  switch) supersedes ⑤ - ⑮ functions.**

### ⑯ Exposure Time Display Window

This window displays the selected exposure time. Estimated air kerma (radiation output) at distal end of cone can be displayed in this window by manual operation or automatically after the exposure. If an abnormal condition exists or a malfunction occurs, an Error Code is also displayed in this window. (See Section: [9] ERROR CODES)

### ⑰ Exposure Warning Light

Illumination of this light indicates the unit is producing x-radiation.

### ⑱ Exposure Switch

This switch initiates radiographic exposure. When making an exposure, depress and hold this switch until the **Exposure Warning Light** ⑰ and the audible warning shut off. Failure to keep this switch depressed will result in the premature termination of the exposure and an error code E.00 will be displayed in **Exposure Time Display Window** ⑯.

## [4] OPERATING PROCEDURES

1. Turn ON the Main Power Switch ①.
2. Confirm that Ready Light ② is illuminated.

**NOTE: The ready light will not illuminate unless the incoming line voltage is correct and within the x-ray's operable range.**

3. Select the appropriate tooth type (⑤ - ⑨), and confirm the pre-selected conditions (cone type, film or digital, kV, mA and patient size) are suitable for exposure.

**NOTE: To manually set the exposure time, depress either of the Manual Exposure Time Adjusting Switches (⤴ or ⤵) until the desired exposure time appears in the Exposure Time Display Window ⑬. While the unit is in manual mode, other selection switches (⑤ - ⑩) do not affect exposure time. (All of the tooth selection lights are off.)**

**To return to the automatic exposure time selection mode, depress any one of Tooth Selection Switches (⑤ - ⑩).**

4. Depress the Exposure Switch ⑭. When the Exposure Switch is depressed, the Exp. Warning Light ⑮ illuminates and the audible warning sounds. Do not release the Exposure Switch until the Exposure Warning Light and audible warning automatically shut off. Failure to keep the switch depressed will result in exposure being terminated prematurely.
5. To continue to radiograph other teeth, just select appropriate Tooth Selection Switches (⑤ - ⑩).

**IMPORTANT: To protect x-ray tubehead from heat accumulation, wait for a time interval that is equal to 50 times the selected exposure time before making additional exposures. (Example : a 25 sec. wait is necessary between exposures that are 0.5 sec. in duration.)**

6. Turn OFF the Main Power Switch ① in order to prevent accidental exposures when the unit is not in use.

**NOTE: If the unit left over 8 min. without being operated and the Main Power Switch ① is kept on, figure "1" runs through the Exposure Time Display Window ⑬. This does not mean that malfunction of the unit has occurred ; this is an energy saving feature. The unit returns to ready condition by pressing any one of the switches, except the Exposure Switch ⑭.**

## [5] ESTIMATED AIR KERMA

Estimated air kerma (radiation output) at distal of cone can be displayed in the exposure time window by depressing the patient switch for more than 1 second. Unit for this value is mGy and this value is calculated by mA, Exposure time and Cone type selected at that time.

Patient type display lamps and displayed value in the window are flashing in this mode and if either of the manual exposure time adjusting switch is depressed during this mode, accumulated air kerma will be displayed. Accumulated value will be reset when the power switch is turned off or leave the x-ray unit more than 8 minutes without depressing any switch. To return to normal mode, press the patient switch for more than 1 second again or leave the controller untouched for more than 10 seconds.

## [6] HAND EXPOSURE SWITCH

Hand exposure switch can be connected to the sub controller. Since this exposure switch has a coiled cord, operators can stand in the most suitable position for operation.

As controller has separate connector for this exposure switch, both exposure switches ⑭ on the front panel of sub controller and this hand exposure switch can be used. If local code prohibits use of two switches, ask installer to disconnect the connector of either switch.

**NOTE : This hand exposure switch is included with FM and FK1/FK2 type x-ray unit.**

## [7] DIGITAL IMAGING SYSTEM

No x-ray image receptor is integrated into the BELRAY II 097 x-ray system. If a receptor for digital imaging is used with BELRAY II 097, the type and performance of the image receptor should be as follows.

1. Type of receptor: CCD (charge-coupled device), CMOS (complementary metal oxide semiconductor) or PSP (photostimulable phosphor plate) receptor for dental intraoral use.
2. Adequate dose of x-radiation for the receptor should be between 0.02mGy and 23.6mGy.
3. Use the receptor holder and receptor cover recommended by the manufacturer of image receptor.
4. Receptor holder should hold the image receptor firmly in position and work as the x-ray beam alignment device.

### WARNING

The use of ACCESSORY equipment not complying with the equivalent safety requirements of the BELRAY II 097 may lead to a reduced level of safety of the resulting system.

Consideration relating to the choice shall include:

- accessory should be CE marked
- evidence that the safety certification of the ACCESSORY has been performed in accordance to the appropriate IEC60601-1 and IEC60601-1 harmonized national standard.

## [8] INFECTION CONTROL AND CLEANING

### 1. INFECTION PREVENTION

X-ray operators are required to wear disposable gloves when taking radiographs and handling contaminated film packets or digital detector cover. Gloves should be changed for each patient to avoid cross contamination. X-ray head, main controller and sub controller should be covered by single use barriers.

### CAUTION

If holders for film or digital detector are used, properly sterilize them according to the procedures indicated by each manufacturer of holders.

### 2. CLEANING

In order to ensure proper hygiene and cleaning of the equipment, the following procedures must be followed.

### CAUTION


Before cleaning the unit, turn off the main power switch and breaker on the branch line. This is required because some internal parts remain connected to main voltage even when the main power switch has been turned off.

Never use the corrosive disinfectants, such as povidone iodine or sodium hypochlorite. Do not pour or spray solvent or liquid directly on the x-ray unit.

Be careful not to allow solvents to run or drip into the x-ray unit.

- a. Turn off the main power switch and breaker on the branch line.
- b. Wipe the outside surface with a paper towel dampened with a disinfectant solution or household, non-abrasive cleaner.  
Recommended disinfectant: FD333 (Durr Dental GmbH)
- c. Allow surface to air dry before turning breaker and main switch back on.

## [9] ERROR CODES

If an abnormal condition exists in the unit, or a malfunction occurs, an error code is displayed in the Exposure Time Display Window . Please refer to the Table below.

Error Code	Condition	Step to be Taken	Possible Solution
E.00	Exposure switch was released before exposure termination.	All the tooth selection lights blink. Depress one of the tooth switches.	Release the exposure switch after the exposure light turns off.
E.01	Exposure switch was depressed within 10 sec. of previous exposure.	A 10 sec. delay is built in between each exposure.	There should be a "wait" interval of 50 times the exposure time between successive exposures.
	Exposure time was set and exposure switch was depressed within 3 sec. of the power switch being turned on.		Wait a minimum 3 sec. after the main power switch is turned on before pressing the exposure switch.
E.02	Line voltage was less than 90% of rated voltage.	Release the exposure switch.	Confirm the ready lamp is ON before exposure. Ask service personnel to check the line voltage.
E.03	Line voltage was more than 110% of rated voltage.		Confirm the ready lamp is ON before exposure. Ask service personnel to check the line voltage.
E.04	Excess current during exposure.	Turn off the main power switch and wait for approximately 2 min. Turn on the main power switch again.	Contact customer service
E.05	Tube current at last portion of exposure was less than 3 mA at 4 mA setting or less than 5.25 mA at 7 mA setting.		If same error code is displayed, call service personnel.
E.06	Tube current at last portion of exposure was more than 5 mA at 4 mA setting or more than 8.75 mA at 7 mA setting.		
E.07	During the exposure, tube current becomes less than 2 mA at 4mA setting or less than 3.5 mA at 7 mA setting.		
E.08	During the exposure, tube current becomes more than 6 mA at 4mA setting or more than 10.5 mA at 7 mA setting.		
E.09	Malfunction of the microcomputer.	Contact customer service	
E.10	Exposure switch or exposure circuit had been ON, when main power switch is turned on.	Release all the switches	Do not turn on the power while other switch is depressed.
E.11	Tube current is detected during pre-heating period.	Turn off the main power switch and wait for approximately 2 min. Turn on the main power switch again.	Contact customer service
E.12	Tube current is detected when main power switch is turned on.		
E.22	Failure of electrical communication between the power PCB and timer PCB.		
E.23	Any switch on the sub controller is depressed when the main power switch is turned on. (Except the exposure switch)	Release all the switches	Do not turn on the power while other switch is depressed.

## [10] MAINTENANCE

BELRAY II 097 x-ray unit requires post installation confirmation and periodic maintenance checks to be performed by dealer service personnel. These procedures ensure that the x-ray unit is functioning within the manufacture’s specifications and remains in compliance with the Standard.

It is responsibility of the owner of the unit to see that these maintenance checks are correctly performed. The specific instructions to perform these checks are located within the BELRAY II 097 Installation instructions.

If it is required by local regulations, make inspectors conduct periodic testing and submit reports as required.

- a. Maintenance personnel: Qualified dealer service personnel who has the experience with Belmont’s x-ray or has been trained by Belmont. But item 6 - 13 of the maintenance check list on page 13 should be verified routinely by treatment room personnel.
- b. Specification of the parameters to be monitored and monitoring frequency: Refer to the maintenance check list on page 13.
- c. Acceptance limit: Refer to the Maintenance check list on page 13.
- d. Required action when failed: Refer to the Maintenance check list on page 13.
- e. Tools to maintain quality control logs: Use the check list on page 13.
- f. Training material: Operator’s instructions, Installation instructions and Service manual.

## [11] DISPOSAL

### 1. Disposal of x-ray unit or components

The tube head of this x-ray unit contains the lead for x-ray shield and oil, which is refined mineral oil and does not contain the carcinogenic substances such as PCBs, for insulation. When disposing the x-ray unit or components, appropriately dispose them complying with all current applicable regulations and local codes. In EU area, EU directive 2012/19/EU on waste electrical and electronic equipment (WEEE) is applied on this product. In this directive, environment conscious recycling / abandonment is obligated.

### 2. Disposal of used film and CCD cover

Dispose the used film covers and CCD sensor covers appropriately, according to procedures indicated by each manufacturer and all current applicable regulations and local codes.

### MAINTENANCE CHECK LIST

Parameter	Acceptance limit	Frequency	Procedures when failed	OK/NG
1. Line voltage	Confirm the line voltage is within 230V±10%. Also confirm the voltage drop during exposure is within 3%.	Yearly	Connect to the power supply within 230V±10%. Check disconnection of wire or connection failure. Repair cable connection as needed.	
2. Tube current	Confirm the measured mA value indicated on the LED window is within the rated value ± 1 mA.	Yearly	Perform MA adjustment. (Refer to installation instructions.)	
3. Mounting plate for wall (WK), ceiling (CK) or floor (FK1/FK2)	Confirm the plate is firmly fixed to the wall (WK), ceiling (CK) or floor (FK1/FK2).	Yearly	If bolts are loose, find the reason why bolts became loose and take counter measure that prevents bolts become loose.	
4. Arm mounting bracket (WK)	Make sure that the arm bracket is firmly attached to the wall or wall plate.	Yearly		
5. Pole (FK1/FK2, CK)	Make sure the pole is securely attached to the mounting plate.	Yearly		

Parameter	Acceptance limit	Frequency	Procedures when failed	OK/NG
6. Dosimetry	Save the image that was taken under appropriate conditions as a reference image. Compare a newly taken image with a reference image to assure the image quality.	Weekly	If the image quality is found poor comparing to a reference image, check the condition of image receptor (film, sensor or imaging plate), image developer (developing fluid, dental film developer, PC or scanner).	
7. Horizontal arm (WK, FK1/ FK2)	Confirm that horizontal arm is firmly inserted to the arm bracket. Make sure the retaining bolt is firmly inserted to the arm bracket.	Daily (before use)	If the retaining bolt is loose, find the reason why bolt became loose, take counter measure that prevent the retaining bolt become loose.	
8. Head	Confirm the head can be smoothly positioned.	Daily (before use)	Adjust the brake screws by referring to installation instructions.	
9. Vertical movement of balance arm	Confirm the balance arm moves smoothly without making noise.	Daily (before use)	Adjust the tension of the balance arm by referring to installation instructions. If the balance arm makes noise, apply grease.	
10. Swing angle of balance arm (FM)	Confirm the balance arm swings between two long legs.	Daily (before use)	Check the stopper screws and mounting screws of pole bushing.	
11. Caster (FM)	Confirm all casters move smoothly and lock function works fine by two lock casters.	Daily (before use)	Clean up the casters or replace them.	
12. Sliding post (RK)	Confirm the post slides smoothly.	Daily (before use)	Check the rollers of sliding post.	
13. Swing arm (CK, RK)	Confirm the joints of the swing arms are connected firmly and stopper and friction are adequate.	Daily (before use)	Check the keys, stopper ring, stopper screws and brake screw of swing arm, and change them as necessary.	



## [12] TECHNICAL DATA


1. X-ray tube-----Canon D-0712 (Stationary Anode)
  - a. Nominal focal spot value-----0.7 (IEC60366)
  - b. Target Material-----Tungsten
  - c. Target angle-----16 deg
  - d. Maximum anode heat content -----4.3 kJ (6 kHU)
2. Maximum x-ray tube assembly heat content -----150 kJ (210 kHU)
3. Rated peak tube potential-----70 kV
4. Rated tube current -----4 mA / 7 mA selectable
5. Maximum rated peak tube potential-----70 kV
6. Rated line voltage -----230 VAC, 50/60 Hz, Single phase
7. Line voltage range-----207 VAC - 253 VAC
8. Range of line voltage regulation-----0 - 2 % (Apparent resistance 1.33 ohm max)
9. Rated line current -----3.4 A at 70 kV (peak), 7 mA
10. Maximum line current -----3.7 A at 70 kV (peak), 7 mA
11. Exposure time -----0.02 - 3.2 sec. (ON and OFF are zero crossed.)
12. Inherent filtration-----1.7 mm Al Equivalent
13. Added filtration -----0.5 mm Al
14. Minimum filtration permanently in useful beam-----2.2 mm Al Equivalent at 70 kV (peak)
15. Nominal radiation output -----Refer to Nominal Radiation Output Table on the next page.
16. Nominal electrical output of H.V. generator -----0.36 kW at 70 kV, 7 mA
17. Cone
 

	Source to skin distance	Field size
a. Regular cone -----	204 mm	58 mm dia., circular
b. Long cone (option) -----	305 mm	58 mm dia., circular
c. Rectangular collimator (option)-----	SSD of cone + 40mm	32 x 40 mm, rectangular
18. Maximum symmetrical radiation field-----60 mm dia. at distal end of cone
19. Leaking technique factor -----70 kV (peak) / 0.137 mA (494 mAs at 1 hour)  
(0.137 mA is maximum rated continuous current for 7mA with a duty cycle 1:50)
20. Duty cycle-----1 : 50 (0.5 sec. exposure with 25 sec. interval)
21. Maximum deviation of tube potential, tube current and exposure time
  - a. Below 0.1 sec. setting ----- $\pm 10$  kV,  $\pm 2$  mA,  $\pm 1$  pulse
  - b. 0.1 sec. setting & up----- $\pm 7$  kV,  $\pm 1$  mA,  $\pm 1$  pulse
22. Measurement base of technique factors
  - a. peak tube potential-----Peak tube potential of conducting half cycle
  - b. tube current-----Average of tube current during one cycle of line frequency
  - c. exposure time-----Impulses of power line frequency
23. Half value layer -----1.5 mm Al over
24. Source to the base of cone distance-----81 mm
25. Environmental condition for storage----- $-20 - 70$  °C, 10 - 99 %, 500 - 1060 hPa
26. Environmental condition for operation----- $10 - 40$  °C, 30 - 75 %, 700 - 1060 hPa
27. Dose area product -----Estimated air kerma displayed [mGy]  
x 26.4 [cm<sup>2</sup>] (for regular and long cone)  
Estimated air kerma displayed [mGy]  
x 12.8 [cm<sup>2</sup>] (for rectangular collimator)
28. Service life-----10 years

## Nominal Radiation Output Table

Exp. Time [sec.]	Nominal Radiation Output															
	without Rectangular Collimator								with Rectangular Collimator							
	50Hz				60Hz				50Hz				60Hz			
	Regular Cone		Long Cone		Regular Cone		Long Cone		Regular Cone		Long Cone		Regular Cone		Long Cone	
	4mA	7mA	4mA	7mA	4mA	7mA	4mA	7mA	4mA	7mA	4mA	7mA	4mA	7mA	4mA	7mA
0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00
0.02	0.08	0.14	0.04	0.07	0.07	0.12	0.03	0.06	0.06	0.10	0.03	0.05	0.05	0.08	0.02	0.05
0.03	0.17	0.28	0.08	0.13	0.14	0.24	0.06	0.11	0.12	0.19	0.06	0.10	0.10	0.17	0.05	0.09
0.04	0.17	0.28	0.08	0.13	0.14	0.24	0.06	0.11	0.12	0.19	0.06	0.10	0.10	0.17	0.05	0.09
0.05	0.25	0.43	0.11	0.20	0.21	0.36	0.10	0.17	0.17	0.30	0.09	0.16	0.15	0.25	0.08	0.13
0.06	0.25	0.43	0.11	0.20	0.28	0.47	0.13	0.22	0.17	0.30	0.09	0.16	0.19	0.33	0.10	0.17
0.08	0.34	0.57	0.15	0.26	0.35	0.59	0.16	0.28	0.24	0.40	0.12	0.20	0.24	0.41	0.13	0.22
0.10	0.42	0.71	0.19	0.33	0.42	0.71	0.19	0.33	0.29	0.49	0.15	0.26	0.29	0.49	0.15	0.26
0.13	0.59	0.99	0.27	0.46	0.56	0.95	0.25	0.44	0.41	0.69	0.21	0.36	0.39	0.66	0.20	0.34
0.16	0.67	1.14	0.30	0.53	0.70	1.18	0.32	0.55	0.46	0.79	0.23	0.41	0.49	0.82	0.25	0.43
0.20	0.84	1.42	0.38	0.66	0.84	1.42	0.38	0.66	0.58	0.98	0.30	0.52	0.58	0.98	0.30	0.52
0.25	1.09	1.85	0.49	0.86	1.05	1.78	0.48	0.83	0.76	1.28	0.38	0.67	0.73	1.23	0.38	0.65
0.32	1.34	2.27	0.61	1.06	1.40	2.37	0.63	1.10	0.93	1.57	0.48	0.83	0.97	1.64	0.49	0.86
0.40	1.68	2.84	0.76	1.32	1.68	2.84	0.76	1.32	1.16	1.97	0.59	1.03	1.16	1.97	0.59	1.03
0.50	2.10	3.55	0.95	1.65	2.10	3.55	0.95	1.65	1.46	2.46	0.74	1.29	1.46	2.46	0.74	1.29
0.63	2.69	4.54	1.22	2.11	2.66	4.50	1.20	2.09	1.87	3.15	0.95	1.65	1.84	3.12	0.94	1.63
0.80	3.36	5.68	1.52	2.64	3.36	5.68	1.52	2.64	2.33	3.94	1.19	2.06	2.33	3.94	1.19	2.06
1.00	4.20	7.10	1.90	3.30	4.20	7.10	1.90	3.30	2.91	4.92	1.48	2.58	2.91	4.92	1.48	2.58
1.25	5.29	8.95	2.39	4.16	5.25	8.88	2.38	4.13	3.67	6.21	1.87	3.25	3.64	6.16	1.86	3.23
1.60	6.72	11.4	3.04	5.28	6.72	11.4	3.04	5.28	4.66	7.90	2.38	4.13	4.66	7.90	2.38	4.13
2.00	8.40	14.2	3.80	6.60	8.40	14.2	3.80	6.60	5.82	9.85	2.97	5.16	5.82	9.85	2.97	5.16
2.50	10.5	17.8	4.75	8.25	10.5	17.8	4.75	8.25	7.28	12.3	3.71	6.45	7.28	12.3	3.71	6.45
3.20	13.4	22.7	6.08	10.6	13.4	22.7	6.08	10.6	9.29	15.7	4.75	8.28	9.29	15.7	4.75	8.28

unit: [mGy]  $\pm$ 50%



## [13] ELECTROMAGNETIC COMPATIBILITY (EMC)

This product conforms to EMC standard EN 60601-1-2:2015+A1:2021.

### 1. Caution to EMC and Compliance with information in attached document

Medical electrical equipment requires special attention to EMC and it must be installed and used according to the EMC information provided in this instruction manual. Do not install in the vicinity of the electrosurgical device being output or electromagnetically shielded room of ME system for MRI diagnostic imaging because the electromagnetic interference intensity is high.

### **WARNING**

- a. Use of this equipment adjacent to or stocked with other equipment should be avoided because it could result in improper operation. If such use is necessary, this equipment and the other equipment should be observed to verify that they are operating normally.**
- b. Use of accessories, transducers and cables other than those specified or provided by the manufacturer of this equipment could result in increased electromagnetic emissions or decreased electromagnetic immunity of this equipment and result in improper operation.**
- c. Portable RF communications equipment (including peripherals such as antenna cables and external antennas) should be used no closer than 30cm (12 inches) to any part of the BELRAY II 097, including cables specified by the manufacturer. Otherwise, degradation of the performance of this equipment could result.**

### 2. Electromagnetic emissions

Emissions test	Test procedure	Compliance	Note:
Conducted and radiated RF emissions	CISPR11	Group 1 Class A	The emissions characteristics of this equipment make it suitable for use in industrial areas and hospitals (CISPR 11 class A). If it is used in a residential environment (for which CISPR 11 class B is normally required) this equipment might not offer adequate protection to radio-frequency communication services. The user might need to take mitigation measures, such as relocating or re-orienting the equipment.
Harmonic distortion	EN 61000-3-2	Class A	
Voltage fluctuations and flicker	EN 61000-3-3	Clause 5	

### 3. Electromagnetic immunity

Immunity test	EN 60601 test level	Compliance level	Electromagnetic environment - guidance
Electrostatic discharge (ESD) EN 61000-4-2	±8 kV contact ±15 kV air	±8 kV contact ±15 kV air	Floors should be wood, concrete or ceramic tile. If floors are covered with synthetic material, the relative humidity should be at least 30%.
Electrical fast transient/burst EN 61000-4-4	±2 kV for power supply lines ±1 kV for input/output lines	±2 kV for power supply lines ±1 kV for input/output lines	Mains power quality should be that of a typical commercial or hospital environment.
Surge EN 61000-4-5	±1 kV differential mode ±2 kV common mode	±1 kV differential mode ±2 kV common mode	Mains power quality should be that of a typical commercial or hospital environment.
Power frequency (50/60Hz) magnetic field EN 61000-4-8	30 A/m	30 A/m	Power frequency magnetic fields should be at levels characteristic of a typical location in a typical commercial or hospital environment.
Proximity magnetic field EN 61000-4-39	134.2kHz 65A/m, Pulse Modulation 2.1kHz 13.56MHz 7.5A/m, Pulse Modulation 50kHz	134.2kHz 65A/m, Pulse Modulation 2.1kHz 13.56MHz 7.5A/m, Pulse Modulation 50kHz	Proximity magnetic fields should be at levels characteristic of a typical location in a professional healthcare facility environment.


Immunity test	EN 60601 test level	Compliance level	Electromagnetic environment - guidance
Voltage dips, short interruptions and voltage variations on power supply input lines EN 61000-4-11	<u>dips</u> 0 %Ut: 0.5 cycle (0, 45, 90, 135, 180, 225, 270 and 315 degree) 0 %Ut: 1 cycle (0 degree) 70 %Ut: 25/30 cycles (0 degree) <u>short interruptions</u> 0 %Ut: 250/300 cycles Ut: Rated voltage of EUT	<u>dips</u> 0 %Ut: 0.5 cycle (0, 45, 90, 135, 180, 225, 270 and 315 degree) 0 %Ut: 1 cycle (0 degree) 70 %Ut: 25/30 cycles (0 degree) <u>short interruptions</u> 0 %Ut: 250/300 cycles Ut: Rated voltage of EUT	Mains power quality should be that of a typical commercial or hospital environment. If the user of the BELRAY II MODEL 097 x-ray requires continued operation during power mains interruptions, it is recommended that the BELRAY II MODEL 097 x-ray be powered from an uninterruptible power supply or a battery.
Conducted RF EN 61000-4-6	AC/DC power and Signal input/output 0.15 MHz - 80 MHz: 3V 6 V in ISM bands between 0.15 MHz - 80 MHz (unmodulated, r.m.s.) 80 % AM (1 kHz)	AC/DC power and Signal input/output 0.15 MHz - 80 MHz: 3V 6 V in ISM bands between 0.15 MHz - 80 MHz (unmodulated, r.m.s.) 80 % AM (1 kHz)	Portable RF communications equipment (including peripherals such as antenna cables and external antennas) should be used no closer than 30cm (12 inches) to any part of the BELRAY II MODEL 097, including cables specified by the manufacturer. Otherwise, degradation of the performance of this equipment could result.
Radiated RF EN 61000-4-3	80 MHz - 2700 MHz: 3V/m (unmodulated, r.m.s.) 80% AM (1kHz)	80 MHz - 2700 MHz: 3V/m (unmodulated, r.m.s.) 80 % AM (1kHz)	
Proximity fields from RF wireless communication equipment EN 61000-4-3	385 MHz 27 V/m (unmodulated, r.m.s.) Pulse modulation 18 Hz	385 MHz 27 V/m (unmodulated, r.m.s.) Pulse modulation 18 Hz	
	450 MHz 28 V/m (unmodulated, r.m.s.) FM $\pm$ 5 kHz deviation 1 kHz sine or Pulse modulation 18 Hz	450 MHz 28 V/m (unmodulated, r.m.s.) FM $\pm$ 5 kHz deviation 1 kHz sine or Pulse modulation 18 Hz	
	710 MHz, 745 MHz, 780 MHz 9 V/m (unmodulated, r.m.s.) Pulse modulation 217 Hz	710 MHz, 745 MHz, 780 MHz 9 V/m (unmodulated, r.m.s.) Pulse modulation 217 Hz	
	810 MHz, 870 MHz, 930 MHz 28 V/m (unmodulated, r.m.s.) Pulse modulation 18 Hz	810 MHz, 870 MHz, 930 MHz 28 V/m (unmodulated, r.m.s.) Pulse modulation 18 Hz	
	1720 MHz, 1845 MHz, 1970 MHz 28 V/m (unmodulated, r.m.s.) Pulse modulation 217 Hz	1720 MHz, 1845 MHz, 1970 MHz 28 V/m (unmodulated, r.m.s.) Pulse modulation 217 Hz	
2450 MHz 28 V/m (unmodulated, r.m.s.) Pulse modulation 217 Hz	2450 MHz 28 V/m (unmodulated, r.m.s.) Pulse modulation 217 Hz		
5240 MHz, 5500 MHz, 5785 MHz 9 V/m (unmodulated, r.m.s.) Pulse modulation 217 Hz	5240 MHz, 5500 MHz, 5785 MHz 9 V/m (unmodulated, r.m.s.) Pulse modulation 217 Hz		

#### 4. Essential performance

Unless the exposure switch is pressed, x-ray is not exposed.



If the Essential performance is lost or deteriorated, the device may operate inadvertently and may harm the patient, the operator, and the surrounding people.


# [14] LABEL LOCATION




**X-RAY ARM**  
 TYPE DENTAL X-RAY COMP 097-A  
 SN EA21E0001  
  
 Takara Belmont Corp.  
 2-1-1, Higashishinbashi, Chuo-ku, Osaka, Japan  
 EC REP TAKARA COMPANY EUROPE GmbH  
 Berner Strasse 18, 60437 Frankfurt am Main, Germany

(WK, FM, FK and CK Type)




**X-RAY HORIZONTAL ARM**  
 TYPE DENTAL X-RAY COMP 097-HA  
 SN EG21E0001  
  
 Takara Belmont Corp.  
 2-1-1, Higashishinbashi, Chuo-ku, Osaka, Japan  
 EC REP TAKARA COMPANY EUROPE GmbH  
 Berner Strasse 18, 60437 Frankfurt am Main, Germany


(WK and FK Type)



(FM Type)

**WARNING**  
 Keep casters in the lock position, unless moving the equipment. To avoid injury, do not push or lean on the equipment.  
**AVERTISSEMENT**  
 Gardez les roulettes en position verrouillée à moins que vous ayez à déplacer l'appareil. Il y a risque de blessure si vous appuyez ou poussez sur l'appareil.

**X-RAY ROOM MOUNT**  
 TYPE DENTAL X-RAY COMP 097-RK  
 SN EB21E0001  
  
 Takara Belmont Corp.  
 2-1-1, Higashishinbashi, Chuo-ku, Osaka, Japan  
 EC REP TAKARA COMPANY EUROPE GmbH  
 Berner Strasse 18, 60437 Frankfurt am Main, Germany



(RK Type)

## [15] OTHER INFORMATION

1. The nature of the emitted radiation;

It is ionizing radiation for diagnostic purposes. Diagnosis is made by obtaining the absorption pattern of this radiation passing through the teeth, jaws, and structures of the oral cavity.

2. The type of the emitted radiation; X-ray

3. Ways of avoiding misuse and of appropriately reducing the risks inherent to transport, storage, and installation;

The pictograms are employed to make the appropriate dose for each tooth and patient size rather than having the operator manually set the exposure time. For transportation and storage, the environment is specified on the outer packaging and in the manual. After installation, the installer is instructed to verify that the device functions correctly.

4. The intensity of the emitted radiation;

Refer to the table on page 15 for the rated dose at the distal end of the cone.

5. The distribution of the emitted radiation;

The radiation area is the infinite conical space (when a round cone is used) or the quadrangular pyramidal space (when a rectangular cone is used) obtained by connecting the boundary of the radiation field at the cone tip and the focal point. The radiation dose is inversely proportional to the square of the distance from the focal point.

6. The recommended dose of emitted radiation;

The dose should be varied depending on the sensitivity of the x-ray receiver used, the teeth to be imaged, and the patient's size. For example, in the case of the Kodak InSight film, the speed setting of F.05 is recommended, and the exposure conditions for each tooth and patient size are shown in the table on page 7. The dose when irradiated under the conditions in this table is shown in the rated dose table on page 15.

7. Means of protecting the patients, the user, or a third party from unintended radiation during the use of medical devices; See [ 1 ] 6.b. on page 1.

## NOTE

EC REP

**TAKARA COMPANY EUROPE GmbH**

Berner Strasse 18, 60437

Frankfurt am Main, Germany

Tel: +49 69 506878 0

Fax: +49 69 506878 20

 **Belmont<sup>®</sup>**



**TAKARA BELMONT CORPORATION**

2-1-1, Higashishinsaibashi, Chuo-ku, Osaka, 542-0083, Japan

TEL: +81 6 6213 5945

TELEFAX: +81 6 6212 3680

Book No. 1A07PUT0

2024-06